



























**Prosthetic components**

<p>Plastic Temporary Abutment                  Plastic Temporary Coping                  Healing Abutment (incl. Healing Abutment Anatomical PEEK and Slim Healing Abutment)                  On1 (IOS) Healing Cap                  Healing Cap Multi-unit</p>					<p>By hand</p>
<p>Slim Temporary Abutment                  Prosthetic Screws for restorations on Multi-unit                  Abutment level                  Titanium Temporary Coping                  Multi-unit</p>				 <p>15 Ncm</p>	
<p>Ball Abutment                  17° Multi-unit Abutment                  30° Multi-unit Abutment</p>					
<p>Abutments for NobelActive 3.0</p> <p>                  Max 15 Ncm</p>					
<p>All other temporary and final abutments for Nobel Blocare implant systems (incl. Temporary Abutment Anatomical PEEK)</p>					 <p>35 Ncm</p>
<p>On1 temporary and final components</p>					

Note: Prosthetic components for other implant systems may require different torque values. Always consult the respective instructions for use.